



ANALYTICAL SUMMARY REPORT

July 14, 2025

Linkan Engineering
2720 Ruby Vista Dr Ste 101
Elko, NV 89801-4943

Work Order: B25070213 Quote ID: B17287

Project Name: Schwartzwalder Mine

Energy Laboratories Inc Billings MT received the following 3 samples for Linkan Engineering on 7/2/2025 for analysis.

Lab ID	Client Sample ID	Collect Date	Receive Date	Matrix	Test
B25070213-001	Outfall 001A	06/27/25 14:15	07/02/25	Aqueous	Solids, Total Suspended
B25070213-002	Outfall 001A	06/30/25 14:00	07/02/25	Aqueous	Same As Above
B25070213-003	Outfall 001A	07/01/25 14:10	07/02/25	Aqueous	Chemical Oxygen Demand Preparation for COD testing HACH 8000 Solids, Total Suspended

The analyses presented in this report were performed by Energy Laboratories, Inc., 1120 So. 27th Street, Billings, MT 59101, unless otherwise noted. Any exceptions or problems with the analyses are noted in the report package. Any issues encountered during sample receipt are documented in the Work Order Receipt Checklist.

The results as reported relate only to the item(s) submitted for testing. This report shall be used or copied only in its entirety. Energy Laboratories, Inc. is not responsible for the consequences arising from the use of a partial report.

Energy Laboratories, Inc. verifies the reported results for the analysis has been technically reviewed and approved for release.

If you have any questions regarding these test results, please contact your Project Manager.



LABORATORY ANALYTICAL REPORT

Prepared by Billings, MT Branch

Client: Linkan Engineering
Project: Schwartzwalder Mine
Lab ID: B25070213-001
Client Sample ID: Outfall 001A

Report Date: 07/14/25
Collection Date: 06/27/25 14:15
Date Received: 07/02/25
Matrix: Aqueous

Analyses	Result	Units	Qualifiers	RL	MCL/ QCL	Method	Analysis Date / By
PHYSICAL PROPERTIES							
Solids, Total Suspended TSS @ 105 C	ND	mg/L		10		A2540 D	07/03/25 09:34 / pjw

Report Definitions: RL - Analyte Reporting Limit
QCL - Quality Control Limit

MCL - Maximum Contaminant Level
ND - Not detected at the Reporting Limit (RL)



LABORATORY ANALYTICAL REPORT

Prepared by Billings, MT Branch

Client: Linkan Engineering
Project: Schwartzwalder Mine
Lab ID: B25070213-002
Client Sample ID: Outfall 001A

Report Date: 07/14/25
Collection Date: 06/30/25 14:00
Date Received: 07/02/25
Matrix: Aqueous

Analyses	Result	Units	Qualifiers	RL	MCL/ QCL	Method	Analysis Date / By
PHYSICAL PROPERTIES							
Solids, Total Suspended TSS @ 105 C	ND	mg/L		10		A2540 D	07/03/25 09:34 / pjw

Report Definitions: RL - Analyte Reporting Limit
QCL - Quality Control Limit

MCL - Maximum Contaminant Level
ND - Not detected at the Reporting Limit (RL)



LABORATORY ANALYTICAL REPORT

Prepared by Billings, MT Branch

Client: Linkan Engineering
Project: Schwartzwalder Mine
Lab ID: B25070213-003
Client Sample ID: Outfall 001A

Report Date: 07/14/25
Collection Date: 07/01/25 14:10
Date Received: 07/02/25
Matrix: Aqueous

Analyses	Result	Units	Qualifiers	RL	MCL/ QCL	Method	Analysis Date / By
PHYSICAL PROPERTIES							
Solids, Total Suspended TSS @ 105 C	ND	mg/L		10		A2540 D	07/03/25 09:34 / pjw
AGGREGATE ORGANICS							
Oxygen Demand, Chemical (COD)	ND	mg/L		5		E410.4	07/03/25 13:58 / fap

Report
Definitions: RL - Analyte Reporting Limit
QCL - Quality Control Limit

MCL - Maximum Contaminant Level
ND - Not detected at the Reporting Limit (RL)



QA/QC Summary Report

Prepared by Billings, MT Branch

Work Order: B25070213

Report Date: 07/14/25

Analyte	Count	Result	Units	RL	%REC	Low Limit	High Limit	RPD	RPDLimit	Qual
Method: A2540 D								Batch: TSS20250703A		
Lab ID: MBLK_20250703-3	Method Blank					Run: BAL #30_250703B		07/03/25 09:34		
Solids, Total Suspended TSS @ 105 C		ND	mg/L	0.6						
Lab ID: LCS_20250703-1	Laboratory Control Sample					Run: BAL #30_250703B		07/03/25 09:34		
Solids, Total Suspended TSS @ 105 C		101	mg/L	25	101	80	120			
Lab ID: B25070196-001CDUP	Sample Duplicate					Run: BAL #30_250703B		07/03/25 09:34		
Solids, Total Suspended TSS @ 105 C		75.8	mg/L	10				4.6	10	
Lab ID: B25070216-005BDUP	Sample Duplicate					Run: BAL #30_250703B		07/03/25 09:34		
Solids, Total Suspended TSS @ 105 C		85.0	mg/L	12				7.4	10	

Qualifiers:

RL - Analyte Reporting Limit

ND - Not detected at the Reporting Limit (RL)



QA/QC Summary Report

Prepared by Billings, MT Branch

Work Order: B25070213

Report Date: 07/14/25

Analyte	Count	Result	Units	RL	%REC	Low Limit	High Limit	RPD	RPDLimit	Qual
Method: E410.4								Analytical Run: SPEC3_250703B		
Lab ID: CCV-201171	Continuing Calibration Verification Standard									07/03/25 13:58
Oxygen Demand, Chemical (COD)		49.2	mg/L	5.0	98	90	110			
Method: E410.4								Batch: 201171		
Lab ID: MB-201171	Method Blank									07/03/25 13:58
Oxygen Demand, Chemical (COD)		ND	mg/L	3				Run: SPEC3_250703B		
Lab ID: LCS-201171	Laboratory Control Sample									07/03/25 13:58
Oxygen Demand, Chemical (COD)		23.1	mg/L	5.0	95	90	110			
Lab ID: B25070255-001CMS	Sample Matrix Spike									07/03/25 13:58
Oxygen Demand, Chemical (COD)		25.6	mg/L	5.0	105	90	110			
Lab ID: B25070255-001CMSD	Sample Matrix Spike Duplicate									07/03/25 13:58
Oxygen Demand, Chemical (COD)		25.3	mg/L	5.0	104	90	110	1.3	10	

Qualifiers:

RL - Analyte Reporting Limit

ND - Not detected at the Reporting Limit (RL)



Work Order Receipt Checklist

Linkan Engineering

B25070213

Login completed by: Natasha L. Anthony

Date Received: 7/2/2025

Reviewed by: ysmith

Received by: DNH

Reviewed Date: 7/11/2025

Carrier name: Return-FedEx NDA

Shipping container/cooler in good condition?	Yes ☒	No ☐	Not Present ☐
Custody seals intact on all shipping container(s)/cooler(s)?	Yes ☒	No ☐	Not Present ☐
Custody seals intact on all sample bottles?	Yes ☐	No ☐	Not Present ☒
Chain of custody present?	Yes ☒	No ☐	
Chain of custody signed when relinquished and received?	Yes ☒	No ☐	
Chain of custody agrees with sample labels?	Yes ☒	No ☐	
Samples in proper container/bottle?	Yes ☒	No ☐	
Sample containers intact?	Yes ☒	No ☐	
Sufficient sample volume for indicated test?	Yes ☒	No ☐	
All samples received within holding time? (Exclude analyses that are considered field parameters such as pH, DO, Res Cl, Sulfite, Ferrous Iron, etc.)	Yes ☒	No ☐	
Temp Blank received in all shipping container(s)/cooler(s)?	Yes ☒	No ☐	Not Applicable ☐
Container/Temp Blank temperature:	4.5°C On Ice		
Containers requiring zero headspace have no headspace or bubble that is <6mm (1/4").	Yes ☐	No ☐	No VOA vials submitted ☒
Water - pH acceptable upon receipt?	Yes ☒	No ☐	Not Applicable ☐

Standard Reporting Procedures:

Lab measurement of analytes considered field parameters that require analysis within 15 minutes of sampling such as pH, Dissolved Oxygen and Residual Chlorine, are qualified as being analyzed outside of recommended holding time.

Solid/soil samples are reported on a wet weight basis (as received) unless specifically indicated. If moisture corrected, data units are typically noted as –dry. For agricultural and mining soil parameters/characteristics, all samples are dried and ground prior to sample analysis.

The reference date for Radon analysis is the sample collection date. The reference date for all other Radiochemical analyses is the analysis date. Radiochemical precision results represent a 2-sigma Total Measurement Uncertainty.

For methods that require zero headspace or require preservation check at the time of analysis due to potential interference, the pH is verified at analysis. Nonconforming sample pH is documented as part of the analysis and included in the sample analysis comments.

Trip Blanks and/or Blind Duplicate samples are assigned the earliest collection time for the associated requested analysis in order to evaluate the holding time unless specifically indicated.

Contact and Corrective Action Comments:

The chain of custody does not indicate which sample is to be analyzed for chemical oxygen demand analysis. Proceeded with chemical oxygen demand on the sample collected 07/01/25 per the sample containers received. YES



Work Order Receipt Checklist - Continued

Linkan Engineering

B25070213

07/11/25

Laboratory Certifications and Accreditations

Current certificates are available at www.energylab.com website:

	Agency	Number
Billings, MT  	Alaska	17-023
	California	3087
	Colorado	MT00005
	Department of Defense (DoD)/ISO17025	ADE-2588
	Florida (Primary NELAP)	E87668
	Idaho	MT00005
	Louisiana	05079
	Montana	CERT0044
	Nebraska	NE-OS-13-04
	Nevada	NV-C24-00250
	North Dakota	R-007
	National Radon Proficiency	109383-RMP
	Oregon	4184
	South Dakota	ARSD 74:04:07
	Texas	TX-C24-00302
	US EPA Region VIII	Reciprocal
	USDA Soil Permit	P330-20-00170
	Washington	C1039
Casper, WY 	Alaska	20-006
	California	3021
	Colorado	WY00002
	Florida (Primary NELAP)	E87641
	Idaho	WY00002
	Louisiana	05083
	Montana	CERT0002
	Nebraska	NE-OS-08-04
	Nevada	NV-C24-00245
	North Dakota	R-125
	Oregon	WY200001
	South Dakota	WY00002
	Texas	T104704181-23-21
	US EPA Region VIII	WY00002
	USNRC License	49-26846-01
	Washington	C1012
Gillette, WY	US EPA Region VIII	WY00006
Helena, MT	Colorado	MT00945
	Montana	CERT0079
	Nevada	NV-C24-00119
	US EPA Region VIII	Reciprocal
	USDA Soil Permit	P330-20-00090



Chain of Custody & Analytical Request Record

Trust our People. Trust our Data.

www.energylab.com

Page 1 of 1

Account Information (Billing Information)

Company/Name Linkan	
Contact Chris Prosper	
Phone 775-777-8003	
Mailing Address 2720 Ruby Vista Dr	
City, State, Zip Elko, NV 89801	
Email AP@linkan.com	
Receive Invoice ☐ Hard Copy ☒ Email	Receive Report ☐ Hard Copy ☒ Email
Purchase Order 25-0152	Quote H17287
	Bottle Order 136995

Report Information (if different than Account Information)

Company/Name Linkan	
Contact Alex Schwiebert	
Phone 775-397-6779	
Mailing Address 2720 Ruby Vista Dr	
City, State, Zip Elko, NV 89801	
Email see comments	
Receive Report ☐ Hard Copy ☒ Email	
Special Report/Format: ☐ LEVEL IV ☐ NELAC ☒ EDD/EDT (contact laboratory) ☐ Other	

Comments

Outfall 001A - Weekly Sample
Please email Report and EDD results to: chris.prosper@linkan.com adam.billin@linkan.com alex.schwiebert@linkan.com peter.hays@state.co.us

Project Information

Project Name, PWSID, Permit, etc. Schwartzwalder Mine	
Sampler Name Robert Acubdo	Sampler Phone 775-7238-6169
Sample Origin State Colorado	EPA/State Compliance ☒ Yes ☐ No
URANIUM MINING CLIENTS MUST indicate sample type	
☐ Unprocessed Ore	
☐ Processed Ore (Ground or Refined) **CALL BEFORE SENDING	
☐ 11(e)2 Byproduct Material (Can ONLY be Submitted to ELI Casper Location)	

Matrix Codes

A - Air	Matrix
W - Water	(See Codes Above)
S - Solids	
V - Vegetation	
B - Bioassay	
O - Oil	
DW - Drinking Water	

Analysis Requested

Total Suspended Solids	
Chemical Oxygen Demand	

All turnaround times are standard unless marked as RUSH.
Energy Laboratories
MUST be contacted prior to RUSH sample submittal for charges and scheduling - See Instructions Page

Sample Identification (Name, Location, Interval, etc.)	Collection Date	Time	Number of Containers	Matrix (See Codes Above)	Analysis Requested	See Attached	ELI LAB ID Laboratory Use Only
1 Outfall 001A	6/27/25	1415	1	W		•	915070213
2 Outfall 001A	6/30/25	1400	1	W		•	
3 Outfall 001A	7/1/25	1410	2	W		•	
4							
5							
6							
7							
8							
9							

ELI is REQUIRED to provide preservative traceability. If the preservatives supplied with the bottle order were NOT used, please attach your preservative information with this COC.

Custody Record MUST be signed	Relinquished by (print) Bryant Acubdo	Date/Time 1510	Signature [Signature]	Received by (print) [Signature]	Date/Time 6/27/25 1015	Signature [Signature]
	Relinquished by (print)	Date/Time	Signature	Received by (print)	Date/Time	Signature
Shipped By	Cooler ID(s)	Custody Seals Y N C B	Intact Y N	Receipt Temp °C	Temp Blank Y N	On Ice Y N
LABORATORY USE ONLY			Payment Type	Amount \$	Receipt Number (cash/check only)	
			Cash			
			Check			

In certain circumstances, samples submitted to Energy Laboratories, Inc. may be subcontracted to other certified laboratories in order to complete the analysis requested. This serves as notice of this possibility. All subcontracted data will be clearly notated on your analytical report.



Trust our People. Trust our Data.
www.energylab.com

Billings, MT 406.252.6325 • Casper, WY 307.235.0515 • Gillette, WY 307.686.7175 • Helena, MT 406.442.0711

BOTTLE ORDER 186995



***** This is a recurring bottle order. If you have received this in error please contact your laboratory *****

SHIPPED Linkan Engineering
TO:

To report an issue with this order, view Safety Data Sheets, or let us know how we are doing, scan here or go to energylab.com/contact-us



Contact: Brendan Smith
400 Corporate Circle, Suite H
Golden CO 80401
Phone: (775) 389-5582
Project: Schwartzwalder Mine - Weekly

Order Created by: Yvonna E. Smith
Shipped From: Billings, MT
Ship Date: 9/17/2024
VIA: Ground
Quote Used: 17287

Bottle Size/Type	Bottles Per Samp	Method	Tests	Critical Hold Time	Preservative	Notes	Num of Samp
------------------	------------------	--------	-------	--------------------	--------------	-------	-------------

Outfall 001A Weekly COD

500 mL Plastic	1	E410.4 HACH 8000	Chemical Oxygen Demand Preparation for COD testing HACH 8000		☒ H2SO4		1
----------------	---	---------------------	---	--	---	--	---

Outfall 001A 3 Times Weekly TSS (3 Sets)

1 Liter Plastic Wide Mouth	1	A2540 D	Solids, Total Suspended			Fill to the neck of the container.	1
----------------------------	---	---------	-------------------------	--	--	------------------------------------	---

Comments

☒ HNO3 - Nitric Acid ☒ H2SO4 - Sulfuric Acid ☒ NaOH - Sodium Hydroxide
☒ ZnAc - Zinc Acetate ☒ HCl - Hydrochloric Acid ☒ H3PO4 - Phosphoric Acid

Material Safety Data Sheets(MSDS) Available @ EnergyLab.com ->Services -> MSDS Sheets

Corrosive Chemicals: Nitric, Sulfuric, Phosphoric, Hydrochloric Acids and Sodium Hydroxide. Zinc Acetate is a skin irritant.

Subcontracting of sample analyses to an outside laboratory may be required. If so, Energy Laboratories will utilize its branch laboratories or qualified contract laboratories for this service. Any such laboratories will be indicated within the Laboratory Analytical Report.

We strongly suggest that the samples are shipped the same day as they are collected.

BO#: 186995

1 of 1



B25070213_Schwartzwalder Mine

1 message

Stacie M. Helms <SHelms@energylab.com>

Mon, Jul 14, 2025 at 9:44 AM

To: "Adam.billin@linkan.com" <Adam.billin@linkan.com>, "Alex.schwiebert@linkan.com" <Alex.schwiebert@linkan.com>, "ap@linkan.com" <ap@linkan.com>, "chris.prosper@linkan.com" <chris.prosper@linkan.com>, "Peter.hays@state.co.us" <Peter.hays@state.co.us>

Thank you for choosing Energy Laboratories Inc. for your analytical needs. Attached you will find the completed report for samples received and processed by Energy Laboratories. No hard copy of the report will be mailed, unless it was requested. Please contact one of our project managers by email at billingspm@energylab.com or by calling 406-252-6325 for questions regarding the attached results. Your invoice will be sent separately.

Please rate your experience or report an issue by clicking [HERE](#) or scanning this QR code below.



Please do not reply to this email.

Sincerely,

Energy Laboratories, Inc.

Trust our People. Trust our Data.

Stacie Helms | Administrative Assistant | Billings, MT

O: 406-869-6295 | shelms@energylab.com | www.energylab.com

This transmission may contain confidential information and is for the use of the intended recipient(s).

If you received this in error, please contact the sender and delete this email and all copies.

2 attachments



B25070213-EDD-ELICSV-WITH-HEADER-1.CSV

2K



B25070213-FINAL REPORT-1.PDF

1736K